

REMARKS

The specification has been amended to correct errors of a typographical and grammatical nature. Due to the excessive corrections thereto, applicants submit herewith a Substitute Specification, along with a marked-up copy of the original specification for the Examiner's convenience. Applicants submit that the substitute specification includes no new matter. Therefore, entry of the Substitute Specification is respectfully requested.

Claims 5 and 9-11 have been amended to correct the dependency thereof and the abstract has also been amended to provide an abstract in proper format and correcting errors of a grammatical nature. A copy of the marked up claims and abstract is also enclosed.

Entry of the preliminary amendments and examination of the application is respectfully requested.


To the extent necessary, applicant's petition for an extension of time under 37 CFR 1.136. Please charge any shortage in the fees due in connection with the filing of this paper, including extension of time fees, to Deposit Account No. 01-2135 (520.40710X00) and please credit any excess fees to such deposit account.

Respectfully submitted,



Melvin Kraus
Registration No. 22,466
ANTONELLI, TERRY, STOUT & KRAUS, LLP

MK/DRA/cee
Attachments
(703) 312-6600



BODY PROBE FOR MRI AND MRI APPARATUS

BACKGROUND OF THE INVENTION

Field of the Invention

5 The present invention relates to a body probe for MRI and an MRI apparatus using the probe⁽¹⁾; and, more particularly, ^{the invention relates} to a body probe for MRI, which is easy to use and provides clear images without image degradation, and an MRI apparatus using the probe.

10

Description of the Related Prior Arts

In general, a body probe for MRI comprises two reception coils that are to be positioned above and under a subject to be examined (examinee). A typical body probe for MRI according to the related art has upper and lower coils formed integrally into a cylindrical shape or a cylindrical shape which can be partly opened sideways, or a belt-like shape.

FIGS. 11A and 11B are diagrams for explaining ^{which show} one structural example of a body probe for MRI according to the related art. Referring to the figures, "111" denotes a body probe, "112" denotes a body portion and "113" ^{denotes} (is) a window.

20 The body probe 111 according to the related art exemplified in FIGS. 11A and 11B has a cylindrical shape. ^a As shown in FIG. 11A, the body probe 111 comprises the cylindrical body portion 112 which has a length and internal vertical and horizontal sizes ^{that are} long and large enough to cover the chest of a subject to be

examined. The windows 13 are formed in the top and bottom portions ^{on} [of] both sides of the body portion 112 in order to lighten the ^{overall} [entire] weight of the body probe and permit the position of the subject with respect to the body probe to be checked from ^{the} outside. Though ~~not~~ ^{not} illustrated, reception coils are embedded in the body portion 112 and are connected to an MRI apparatus, so that signals from the coils ^{can be} [are] processed to form an image to be displayed.

FIG. 11B shows how the cylindrical body probe 111 [according to the related art] is used. A subject to be examined, whose chest is enclosed by the body probe 111, is placed inside an MRI apparatus for examination. Generally speaking, the size of ^{a person's} [the] chest varies from one person to another. It is therefore difficult to prepare a probe which ^{precisely} matches [with] the size of the chest of each subject, and such an attempt inevitably results in an increase in the cost of the entire MRI system. ^{For} ^{reason} [In] this [respect], normally, three or so different sizes of body probes 111 are prepared and used for all subjects of different builds.

Although the body probe exemplified in FIGS. 11A and 11B is designed ^{to be} completely cylindrical, there is a body probe designed in such a way that it can be opened sideways at the center portion.

FIG. 12 is a diagram ^{showing} [for explaining] another structural example of ^a [the] body probe for MRI [according to the related art]. In FIG. 12, "121" ^{indicates an opening} [is a bore], "122"

^{denotes} [is] an upper coil, "123" ^{denotes} [is] a lower coil, "124" ^{denotes} [is] a belt, "125" denotes cords, and "126" ^{denotes} [is] a connector box.

The body probe for MRI exemplified in FIG. 12 has reception coils, which comprise the upper coil 122 and the lower coil 123. In using the body probe, the lower coil 123 and the upper coil 122 are secured by the belt 124. ^{Lower coil is} The ~~(former)~~ placed under the back of a subject to be examined and the ^{upper coil is placed} ~~(latter)~~ over the abdominal portion, the chest or the like of the subject. The belt 124 is attached to a belt groove provided in the frame of a bed in such a way as to be movable along the bed to secure ^{the subject at} a desired portion ~~[of the subject]~~. The belt 124 is movable in accordance with the mounting position of the upper coil 122. The body probe for MRI [^] [according to the related art] which has the above-described structure can securely fix the upper coil 122 and the subject together.

The cords 125 that extend from the upper and lower coils run along one edge portion of the bed and are connected to the connector box 126 located inside the ^{opening} ~~(bore)~~ 121 of the MRI apparatus. At the time of picking an image with the MRI apparatus, the subject is ^{moved} ~~(fed)~~ into the ^{opening} ~~(bore)~~ 121 to a predetermined position together with the top table of the bed and the connector box 126.

The cylindrical body probes for MRI ^{as described above} ~~(according to)~~ [^] [the related art] ~~do not~~ not have fixed distances between the subject and the coils in the body probe due to a

difference in the build of the subject, and [has a] difficulty^{is experienced} in securing the subject inside the body probe. This makes the characteristics of the coils unstable, so that the picked-up image of the subject is degraded by [the] movement of the body of the subject. The cylindrical body probe also has a shortcoming [such] in that it is not easily attached to a subject.

Another body probe for MRI [according to the], [related art] whose reception coils comprise an upper coil and a lower coil, has no coil portions provided at portions corresponding to the sides of the subject. In addition, it is difficult to place the center of the upper coil over the center of the body axis of the subject, so that the center of the upper coil may be misaligned with^{respect to} the center of the body axis of the subject. This makes it difficult to acquire clear images.

With regard to (the)^{use of a} belt-like body probe [according to the related art], one among several sizes of body probes which matches [with] the build of a subject to be examined is selected and is fitted on the subject. This type^{of body probe} cannot [suppress the]^{restrain} movement of the body of the subject, so that the body probe moves as the body of the subject moves. This also leads to image degradation.

The reception coils^{of the above-described body probes} [according to the related art] are cylindrical or belt-like and their sizes are not prepared subject by subject. Rather, one among

several sizes of ^{available} body probes ^{previously} prepared, which is closedst to the build of a subject to be examined, is selected. As the shapes of the reception coils become closer to the shape of the subject, the sensitivity becomes higher. To achieve a high sensitivity, therefore, it is necessary to fit the reception coils closely to the subject. On the contrary, the reception coils of the body probes ^{of the type described above} (according to the related art) are difficult to ~~be~~ fit~~ted~~ closely to each subject to be examined.

10

SUMMARY OF THE INVENTION

Accordingly, it is an object of the ^{present} invention to provide an easy-to-use body probe for MRI, which is able to overcome the aforementioned problems [of the related art], is easily attached to various subjects to be examined who have ^{builds of} different sizes [of builds] and does not suffer image degradation, and an MRI apparatus which uses this probe.

15

The ^{foregoing} object is achieved by designing a body probe for MRI having reception coils for picking up an image for use in an MRI apparatus in such a way that the reception coils are provided in a center coil portion and two side coil portions.

20

The ^{foregoing} object is achieved by providing slide mechanisms ^{on} [of] the side coil portions, which are formed by smooth surfaces ^{that} are formed on both right and left sides of the center coil portion, at joint portions between the center coil portion and the side coil

25

portions.

The ^{foregoing} object is achieved by designing a bottom surface of the center coil portion ^{to have} (in) a curved shape, or providing an indicator indicating ^{the} (a) center of (a) ^{the} center coil at ^{the} (a) center of a top surface of the center coil portion. The indicator indicating the center of the center coil may be a groove, a projection or a line drawn in ^a color different from ^{the} (a) color around the line.

The ^{foregoing} object is also achieved by designing a body probe for MRI having reception coils for picking up an image for use in an MRI apparatus in such a way that the reception coils are provided in a center coil portion, side coil portions and a base coil portion. The center coil portion, the side coil portions and the base coil portion may be constructed independently of one another.

The ^{foregoing} object is achieved by constructing the side coil portions by embedding a flexible coil in a soft member, ^{by} or providing the side coil portions with cutaway holes at portions where there are no coil portions.

(The) Other objects and structures of the invention and the resulting advantages ^{there} will become ^{more} apparent from the following detailed description of an embodiment of the invention.

BRIEF DESCRIPTION OF THE DRAWINGS

FIGS. 1A and 1B are perspective views showing the outward appearance of a body probe for MRI

according to one embodiment of the invention;

FIG. 2 is ^{an exploded perspective view} [a diagram] showing side coil portions of the body probe for MRI according to the embodiment of the invention [opened];

5 FIG. 3 is a ^{perspective view illustration} [diagram for explaining] the procedures of fitting the body probe according to the embodiment of the invention to a subject to be examined;

10 FIG. 4 is a ^{perspective view illustration} [diagram for explaining] the state in which the body probe according to the embodiment of the invention is fitted on the subject (examinee);

FIG. 5 is a ^{cross-sectional illustration} [diagram for explaining] how the side coil portions are deformed;

15 FIG. 6 is a ^{perspective view illustration} [diagram for explaining] another means for fitting the center coil portion on the subject;

FIGS. 7A and 7B are diagrams ^{illustration} [for explaining] the basic shapes of coils formed in individual coil portions according to the embodiment of the invention and the shape of a coil formed in a head coil portion;

20 FIG. 8 is a diagram ^{illustration} [for explaining] the shape of a coil formed in a side coil portion;

FIG. 9 is a ^{perspective view illustration} [diagram for explaining] the shape of a coil formed in a base coil portion;

25 FIG. 10 is a ^{perspective view illustration} [diagram for explaining] the shape of a coil formed in a center coil portion and the structure of a cable connector;

FIGS. 11A and 11B are ^{perspective view illustration} [diagrams for explaining]

one structural example of a body probe for MRI according to the related art; and

FIG. 12 is a ^{perspective view illustration} diagram for explaining another structural example of the body probe for MRI according to the related art.

DESCRIPTION OF THE PREFERRED EMBODIMENTS

A preferred embodiment of the invention ^{will be} described below with reference to the accompanying drawings.

FIGS. 1A and 1B are perspective views showing the outward appearance of a body probe for MRI according to one embodiment of the invention, and FIG. 2 is a diagram showing the side coil portions of the

body probe according to the embodiment of the invention, ^{in the} opened ^{condition}. In the ^{drawing} diagram, ^{the body probe has} "11" is a base coil portion, "12" is a side coil portion, "13" is a center coil portion, "14" is a head coil portion, "15" is a head holder, "16" is a fixing band, "17" is a coil cable, "18" and "19" are cutaway holes, "20" is a center indicator portion, "21" is a cable connector, "22" is a fixing-band groove, "23" is a smooth surface, "24" is a head-coil connecting portion, "25" are side-coil connecting portions, "26" is a connector, and "27" is a recess portion.

As shown in FIGS. 1A and 1B (which are perspective views from different viewpoints that will not be distinguished from one another in the following

description) and FIG. 2, the body probe for MRI
according to the embodiment of the invention comprises
— (the) base coil portion 11, which is positioned at the
back portion of a subject to be examined (hereinafter
5 simply referred to as ^{the} "subject") when ^{the person lays down} (laid down) on his
or her back on an unillustrated bed, the two side coil
portions 12, which are to be connected to the side-coil
connecting portions 25 provided on both sides of the
base coil portion 11, and the center coil portion 13,
10 which connects the two side coil portions 12 at the
center of the chest of the subject.

The head holder 15 is provided at that portion
of the base coil portion 11 where the head of the
subject is to be placed. The head holder 15 holds the
15 head of the subject in such a way that the subject's
head does move around. The recess portion 27 where the
buttocks of the subject will rest is provided at the
opposite side of the base coil portion 11 ^{relative} to the head
holder 15. The side-coil connecting portions 25 are
20 formed on both sides of the base coil portion 11, and
the side coil portions 12 are to be attached in such a
way as to be movable at an arbitrary angle with respect
to the base coil portion 11. This design makes it
possible to open the side coil portions 12, as shown in
— FIG. 2, to facilitate attachment of the body probe to
25 the subject at the time of attachment. Further,
folding the side coil portions 12 inside makes it
easier to carry the side coil portions 12 around and

can contribute to reducing the storage space therefor. The side coil portions 12 can be detached from the base coil portion 11 at the positions of the side-coil connecting portions 25.

5 Each side coil portion 12 is made of a soft member, such as a rubber-based or urethane based member, and has a band-like conductive member embedded inside. As will be discussed later, therefore, the side coil portions 12 can take shapes which match (with) the side
10 surfaces of the subject and can be fitted closely on the subject. Even when the subject puts (the) weight on the side coil portions 12, the soft members used for the side coil portions 12 neither cause (a) pain on the subject nor damage ^{to} the side coil portions 12. For the
15 purpose of weight reduction, the cutaway hole 18 is formed in that portion of each side coil portion 12 where the internal coil does not exist. To improve the image quality, that portion of each side coil portion 12 which is to be connected to the center coil portion
20 13 is designed to be arcuately curved in such a way that the curved portion comes closest to the coil center of the base coil portion 11.

 Two coil cables 17 having connectors 26 are provided on each side of those portions of the two side
25 coil portions 12 which are to be connected to the center coil portion 13 at the chest of the subject. Provided at the center portions of those portions of the two side coil portions 12 which are to be connected

to the center coil portion 13 at the chest of the subject are the fixing bands 16 for fixing the side coil portions 12 and the center coil portion 13 in close contact with the subject. A cutaway hole 18' is formed in each side coil portion 12 on ^{the} ~~(that)~~ side ^{close to} ~~(of)~~ the base coil portion 11. When the side coil portions 12 are fitted on the subject, portions of the subject can be seen through the cutaway holes 18'. The cutaway holes 18' also serve to improve the ventilation.

The center coil portion 13, which is to be placed at the center of the chest of the subject, has capabilities to couple the two side coil portions 12 together and electrically connect the coils in the coil portions. The center coil portion 13 improves the close attachment to the subject and always stays on the center of the body axis of the subject, so that the constant precision of picked-up images can be maintained. To ensure adequate coil shapes at the time the center coil portion 13 is connected to the side coil portions 12, the bottom surface of the center coil portion 13 is curved to match the shape of the chest of the subject. This design can reduce the sensation of pressure on the subject.

As shown in FIG. 2, the center coil portion 13 is provided with ~~(the)~~ smooth surfaces 23 on the right and left sides, which constitute slide mechanisms. The center portion of the center coil portion 13 that ^{is adjacent to} ~~(follows)~~ the smooth surfaces 23 is a step portion which

is thicker than the smooth surfaces 23. Provided at the step portion of the center portion are the cable connectors 21 to which the connectors 26 attached to the coil cables 17 of the side coil portions 12 are connected. The center indicator portion 20 that indicates the center is provided at the position of the center portion, which becomes the center of the body axis, as a groove, a projection or a line drawn in, a color different from ^{the} (a) color around the line. Further provided in the center portion of the center coil portion 13 in a direction perpendicular to the body axis is the fixing-band groove 22 through which the fixing bands 16 are ^{led} put.

The head coil portion 14 is suitable for use with the body probe according to the embodiment of the invention. The head coil portion 14, when in use, is coupled to the head-coil connecting portions 24 formed on both sides of the head holder 15. As will be discussed later, the cutaway holes 19 in the head coil portion 14 are provided at the positions of ^{the} ears of the subject ^{as is} to suppress the sensation of pressure at the time the head coil portion 14 is mounted on the subject and to permit smooth conversation with a caretaker.

FIG. 3 is a diagram, ^{illustration} (for explaining) the procedures ^{involved in} (of) fitting the body probe according to the embodiment of the invention on a subject; FIG. 4 is a diagram, ^{illustration} (for explaining) the state in which the body probe according to the embodiment of the invention is

fitted on the subject^{illustration} and FIG. 5 is a diagram (for) (explaining) how the side coil portions are deformed. Referring now to FIGS. 3 to 5, the procedures (of) used in fitting the body probe on the subject and the fitted state ^{will be} (are) described below.

5 With the side coil portions 12 opened sideways and the head coil portion 14 removed, as shown in FIG. 2, the subject is laid down on the base coil portion 11 in such a way as to place the subject's head in a cavity of the head holder 15, as shown in FIG. 3. At this time, the buttocks of the subject are positioned in the recess portion 27 of the base coil portion 11. Next, the caretaker places the center coil portion 13 on the chest of the subject and adjusts its position in such a way that the center indicator portion 20 ^{coincides with} (comes to) the center of the chest of the subject and the fixing-band groove 22 ^{is aligned with} (comes to) the position of the fixing bands 16 of the side coil portions 12.

15 Then, the caretaker turns the two side coil portions 12 inward so as to enclose the chest of the subject and connect the side coil portions 12, ^{using the fixing bands 16, but} without altering the position of the center coil portion 13 (by) (the fixing bands 16). At this time, the distal portions of the side coil portions 12 are placed on the smooth surfaces 23 of the center coil portion 13, and ^{they} can slide on the smooth surfaces 23 ^{in response to} (by) the tension ^{produced by} (of) the fixing bands 16. Accordingly, the side coil portions 12, which are constructed by embedding band-like conductive

members inside soft members, can be deformed to match the shape of the chest of the subject and ^{wrapped} (put) around the chest of the subject tightly, as shown in FIG 4.

Next, the caretaker ^{will} attach the head coil portion 14 to the head-coil connecting portions 24 on the base coil portion 11. As seen from FIG. 4, when the head coil portion 14 is attached, the ears of the subject come to the positions of the cutaway holes 19 formed in the head coil portion 14. Therefore, the subject can hear (the) ambient sounds without interference and can feel relaxed about the examination.

Surface coils to be discussed later are formed inside the side coil portions 12 at portions close to the center coil portion 13 and around the cutaway holes 18, protruding in the direction of the shoulders and the direction of the waist. Those portions of the side coil portions 12 where the surface coils are not located are cut away. The subject can therefore put ^{his or her} (the) arms out through the cutaway portions loosely, thus reducing the awkwardness of putting the probe on. The coils that are formed to protrude in the direction of the shoulders and the direction of the waist can provide an image over a wide range.

Referring now to FIG. 5, a description will be given of how to deform the above-described side coil portions 12 to match the shape of the chest of the subject and ^{wrap} (put) them around the chest of the subject tightly.

As shown in FIG. 5, the stepped center portion of the center coil portion 13 that follows the smooth surfaces 23 thereof has a total width of 60 mm, 30 mm in either direction from the center indicator portion 20, and the smooth surface 23 that constitutes the slide mechanism on one side has a width of 75 mm. The distal end portion of each of the two side coil portions 12 can move within a range of the 75-mm width of the smooth surface 23. Accordingly, the inner circumferential length, that is formed by the base coil portion 11, the two side coil portions 12 and the center coil portion 13, can be adjusted within a range of 150 mm. This state is illustrated in FIG. 5. Therefore, the side coil portions 12 can be deformed to match the size of the chest of the subject and [put] be wrapped around the chest of the subject tightly.

As mentioned above, the body probe according to the embodiment of the invention is designed in such a way that a set of one center coil portion 13 and two side coil portions 12 can absorb a difference of 150 mm in the circumferential length of the chest of the subject and the side coil portions 12 can be deformed and put around the chest of the subject tightly. To absorb a difference in the circumferential length of the chest of a larger subject, another center coil portion 13 of a different width size should be prepared. This can allow the body probe of the invention to cope with all types of subjects from a small-sized Japanese person

to a large-sized European ^{person} ~~conventional~~

In case of the ^{conventional} cylindrical body probe [according] ~~to the related art~~, at least three types of body probes, large, intermediate and small ones, ^{need to} [should] be prepared.

5 By way of contrast, the body probe according to the embodiment of the invention can be used ^{for} all types of subjects of different builds by merely preparing two types of center coil portions 13.

FIG. 6 is a diagram ^{illustrating} [for explaining] another means for fitting the center coil portion on the, ^{the probe includes} subject. In FIG. 6, ¹⁶ "16" denotes a fixing belt, and the other structural elements, which are not given reference symbols, are the same as those explained above with reference to FIGS. 1A, 1B, 2, 3, 4 and 5.

15 The body probe according to the embodiment of the invention that has already been discussed ^{with reference} [referring] to FIGS. 1A, 1B, 2, 3, 4 and 5 is designed in such a way that the side coil portions 12 and the center coil portion 13 are fixed in close contact with a subject by ^{use of} the fixing bands 16 provided on the side coil portions 12. By way of contrast, the example shown in FIG. 6 is designed in such a way that the fixing belt 16' is attached to the sides of the base coil portion 11. The body probe is attached to the subject by first placing ^{using} the center coil portion 13 on the chest of the subject, then securing the center coil portion 13 on the chest of the subject ^{using} [by] the fixing belt 16', and then connecting the side coil portions 12 to the center coil

20

25

portion 13 in the same way as has been discussed in the foregoing description.

According to the example shown in FIG. 6, the side coil portions 12 have only to be connected to the center coil portion 13 after the center coil portion 13 is secured on the chest of the subject ^{using} by the fixing belt 16'. The work of attaching the body probe to the subject becomes easier than the [attachment] work [of] ^{involved in} securing the side coil portions 12 and the center coil portion 13 together in close contact with the subject merely by ^{means of} the fixing bands 16 [involved for] ⁱⁿ the body probe that has already been described [referring] ^{with reference} to FIGS. 1A, 1B, 2, 3, 4 and 5. This feature can [relieve a] ^{reduce the burden on the} caretaker [of the burden].

The foregoing description has ^{been directed to an} [explained the] embodiment of ^a [the] body probe for MRI according to the invention. A description will now be given of the shapes of coils formed in the individual coil portions that have already been discussed.

FIGS. 7A and 7B are diagrams ^{illustration} [for explaining] the basic shapes of coils formed in the individual coil portions according to [the embodiment of] the invention and the shape of the coil formed in the head coil portion. FIG. 8 is a diagram ^{illustration} [for explaining] the shape of the coil formed in the side coil portion, and FIG. 9 is a diagram ^{illustration} [for explaining] the shape of the coil formed in the base coil portion. FIG. 10 is a diagram ^{illustration} [for explaining] the shape of the coil formed in the

center coil portion and the structure of the cable
connector. In the ^{diagram} diagrams, ⁷¹["71"] is a solenoid coil,
["72" is] a surface coil, ⁷³["73" is] a solenoid coil for
cervical vertebrae, ⁷⁴["74" is] a surface coil for head,
5 ["81" is] a solenoid coil for lumbar vertebrae, ⁸²["82" is] a
solenoid coil for thoracic vertebrae, ⁸³["83" is] a surface
coil for lumbar vertebrae, ⁸⁴["84" is] a surface coil for
thoracic vertebrae, and ^{91, 92}["91" and "92" denote] connectors.

According to the invention, as shown in FIG. 7A,
10 the combination of both the solenoid coil and the
surface coil is used in picking up the image of a
subject's part of interest. The solenoid coil 71 is
formed in a loop and is laid out to surround the
subject's part to be picked up. The surface coil 72 is
15 formed in the shape of the letter "8" in such a way as
to be extendible in a planar pattern in the coil
portion. The surface coil 72 is laid along the
subject's part to be picked up.

As shown in FIG. 7B, the cervical-vertebrae
20 solenoid coil 73 for picking up the image of the
cervical vertebrae is provided in the bridge that is
located in the head coil portion 14 at a position close
to the neck of the subject. The cervical-vertebrae
solenoid coil 73 is connected inside the base coil
25 portion 11 via unillustrated connectors provided in the
head-coil connecting portions 24, so that the entire
solenoid coil 73 is formed in a loop. The head surface
coil 74 is laid around each cutaway hole 19 formed in

the head coil portion 14. The head surface coil 74, like the solenoid coil 73, is connected inside the base coil portion 11 via an unillustrated connector provided in the associated head-coil connecting portion 24, so
5 that the entire solenoid coil 73 is formed in the shape of the letter "8". That internal part of the base coil portion 11 which constitutes a part of the head surface coil 74 has a shape ^{that is elongated} [extending long] toward the associated shoulder.

10 As shown in FIG. 8, the lumbar-vertebrae solenoid coil 81, the thoracic-vertebrae solenoid coil 82, the lumbar-vertebrae surface coil 83 and the thoracic-vertebrae surface coil 84 are arranged inside each side coil portion 12. The coils 81 and 82 are
15 used to pick up the images of the lumbar vertebrae and thoracic vertebrae. As in the case of the head coil portion 14, the solenoid coils 81 and 82 are connected inside the base coil portion 11 and the center coil portion 13 to form loops and the surface coils 83 and
20 84 are connected inside the base coil portion 11 and the center coil portion 13 to form the shapes of the letter "8". That side of each coil which is to be connected to the center coil portion 13 is connected to the connector 26 for connection to the center coil
25 portion 13 via the associated coil cable 17, as has already been explained.

As apparent from the foregoing description, the base coil portion 11 connects the coils provided in the

side coil portions 12 and the head coil portion 14 to form ~~(the desired shapes of the)~~ solenoid coils and ~~(the)~~ surface coils. ^{of desired shapes} As shown in FIG. 9, coil portions which form the individual coils provided in the side coil

5 portions 12 are laid inside the base coil portion 11. The connection to each side coil portion 12 is accomplished via the connector 91 provided at the side coil portion 12 and the connector 92 provided at the base coil portion 11. Though not illustrated, the coil
10 portions that form the individual coils in the head coil portion 14 are likewise laid inside the base coil portion 11.

As shown in FIG. 10, ^{two} ~~(the)~~ cable connectors 21 are provided ~~(two)~~ on a single circuit board and are
15 provided in the center coil portion 13 in such a way that two cable connectors 21 are positioned on either side of the center portion of the center coil portion 13 that is designed to have a step thicker than the smooth surfaces 23.

20 Though unillustrated, the individual coils that constitute the body probe and the coils provided in the head coil portion are connected to the body of the MRI apparatus via cables for acquisition of reception signals, so that an image is produced from the
25 reception signals.

According to ~~(the embodiment of)~~ the invention, ^{the provision of} the coils ~~(are provided)~~ in separated plural coil portions and the use of the center coil portion

facilitate the attachment of the coils to a subject in such a way that the center of each coil is always aligned with the center of the body axis of the subject. According to (the embodiment of) the invention, each side
5 coil portion is constructed by embedding a flexible coil in a soft member and the position at which the center coil portion is joined to the side coil portions can be slidably adjusted. It is therefore possible to allow the reception coils to be attached to the body of
10 the subject tightly, ^{to accommodate any} (in accordance with a) difference in the build of the subject. This can provide high-precision images.

According to the embodiment of the invention, because the slide adjustment can allow the reception
15 coils to be attached to the body of the subject tightly, the coil center does not deviate even when the subject moves. As a result, high-precision images can be obtained. ^{Since} (As) the center coil portion 13 has (the) ^a
20 portion that permits slide adjustment of the side coil portions, preparation of two types of center coil portions of different width sizes can allow the body probe to be used for all types of subjects of different builds.

According to the embodiment of the invention, ^{since}
25 cutaway holes are formed in those portions of the individual coil portions where there are no coils laid out and the member for molding the coils is a soft member, such as a rubber-based or urethane based member,

it is easy to fit the body probe on the subject, and it is possible to improve the sensation of attachment and the feeling of freedom.

5 As described above, the invention can provide an easy-to-use body probe for MRI which is easily attached to various types of subjects to be examined and does not suffer image degradation, and an MRI apparatus which uses this probe.

Industrial Applicability

10 A body probe for MRI according to the invention and an MRI apparatus using the probe can be effectively used in [the field of] the medical equipment industry.

WHAT IS CLAIMED IS:

1. A body probe for MRI having reception coils for picking up an image for use in an MRI apparatus, said reception coils being provided in a center coil portion
5 and two side coil portions.

2. The body probe for MRI according to claim 1, wherein slide mechanisms of said side coil portions are provided at joint portions between said center coil portion and said side coil portions.

10 3. The body probe for MRI according to claim 2, wherein smooth surfaces are formed on both right and left sides of said center coil portion and constitute said slide mechanisms.

15 4. The body probe for MRI according to any one of claims 1 to 3, wherein a bottom surface of said center coil portion is curved.

→ 5. The body probe for MRI according to (any one of) ^{claim} (claims) 1 (to 4), wherein an indicator indicating a center of a center coil is provided at a center of a top
20 surface of said center coil portion.

6. The body probe for MRI according to claim 5, wherein said indicator indicating said center of said center coil is a groove, a projection or a line drawn in color different from a color around said line.

25 7. A body probe for MRI having reception coils for picking up an image for use in an MRI apparatus, said reception coils being provided in a center coil portion, side coil portions and a base coil portion.

8. The body probe for MRI according to claim 7, wherein said center coil portion, said side coil portions and said base coil portion are constructed independently of one another.

5 9. The body probe for MRI according to [any one of] claims 1 to ^{3,7 and 8} 8, wherein said side coil portions are constructed by embedding a flexible coil in a soft member.

10 10. The body probe for MRI according to any one of claims 1 to ^{3,7 and 8} 9, wherein said side coil portions are provided with cutaway holes at portions where there are no coil portions.

15 11. An MRI apparatus that uses a body probe which, as recited in any one of claims 1 to ^{3,7 and 8} 10, has reception coils for picking up an image.

ABSTRACT OF THE DISCLOSURE

(Disclosed are an easy-to-use body probe for MRI
[which is easily attached to a subject to be examined]
[and does not suffer image degradation, and an MRI]
5 [apparatus using this probe. The ^Abody probe for MRI
[according to the invention comprises] a base coil
portion 11, two side coil portions 12 to be connected
to side-coil connecting portions 25 provided on both
sides of the base coil portion 11, and a center coil
10 portion 13 which connects the two side coil portions 12
at the center of the chest of the subject. Each side
coil portion 12 is made of a soft member and has a
flexible coil embedded inside, so that the side coil
portion can have a shape which matches [with] the side
15 surface of the subject. Provided at the center
[portions of those portions] of the two side coil
portions 12 which are to be connected to the center
coil portion 13 at the chest of the subject is a fixing
band 16 for fixing the side coil portions 12 and the
20 center coil portion 13 in close contact with the
subject. [The probe structure can provide an easy-to-]
[use body probe for MRI which is easily attached to a]
[subject to be examined and acquires high-quality MRI]
[images.]